



CentriFlow®

Material Test Report

Elbow Macaroni



CentriFlow®

Date Tested:	June 13, 2003	Temperature:	Ambient (78°F/25.6°C)
Technician:	James Seagraves	Particle Size:	1/4" to 1/2" pieces
Test Location:	Eastern Instruments	Flowability:	Good
CFM Model:	12" Type I CentriFlow®	Cohesiveness:	Slight
Meter Capacity:	15 ft³/min	Density (lb/ft³):	45 lb/ft³
Feed System:	Belt Conveyor	Inhibit Setting:	0.200 Volts



Test #1	EFS = 12,000 lb/hr				Mass Flow Rate = 3,000 lb/hr				
Run #	Actual Weight		Metered Weight		Actual/Metered		Delta Weight		% Error
1	8.49		8.50		0.999		0.01		0.12%
2	8.81		8.81		1.000		0.00		0.00%
3	8.20		8.21		0.999		0.01		0.12%
4	7.86		7.88		0.997		0.02		0.25%
Average:					0.999				
STD:					0.00104				
% STD:					0.10%				
<p>Additional Comments: Tested using a 12" CentriFlow® Type I Meter in the In-Line Flow Configuration. The run time was 7 seconds per batch.</p>									

Accuracy Statement:

"The CentriFlow® Meter will provide accuracy to within ±0.25% of reading when operating within ± 10% of the calibrated flow rate, as long as the flow rate is within the operational range of the meter."