



E A S T E R N I N S T R U M E N T S



CentriFlow®

# Material Test Report

Pale Malt



CentriFlow®

Date Tested:	October 24, 2001	Temperature:	Ambient (78°F/25.6°C)
Technician:	Alan Norman	Particle Size:	0.25"
Test Location:	Eastern Instruments	Flowability:	Above Average
CFM Model:	6" Type II CentriFlow®	Cohesiveness:	None
Meter Capacity:	6.75 ft³/min	Density (lb/ft³):	35 lbs/ft³
Feed System:	Screw Conveyor	Inhibit Setting:	0.200 Volts



Test #1		Percent of Volumetric Capacity = 50%					
Run #	Actual Weight	Metered Weight	Actual/Metered	Delta Weight	% Error		
1	30.16	30.11	0.998	-0.05	-0.17%		
2	30.18	30.18	1.000	0.00	0.00%		
3	32.60	32.56	0.999	-0.04	-0.12%		
4	31.76	31.72	0.999	-0.04	-0.13%		
5	33.32	33.32	1.000	0.00	0.00%		
Average:			0.999				
STD:			0.0008				
% STD:			0.08%				
Additional Comments: Tested using a 6" CentriFlow® Type II Meter equipped with VibraWeigh®. The run times were 15 seconds per batch.							

### Accuracy Statement:

"The CentriFlow® Meter will provide accuracy to within  $\pm 0.25\%$  of reading when operating within  $\pm 10\%$  of the calibrated flow rate, as long as the flow rate is within the operational range of the meter."