



EASTERN INSTRUMENTS



CentriFlow®

# Material Test Report

Ground Shingles



CentriFlow®

Date Tested:	May 31, 2012	Temperature:	Ambient (78°F/25.6°C)
Technician:	James Seagraves	Particle Size:	Powder to 0.50" chips
Test Location:	Eastern Instruments	Flowability:	Average
CFM Model:	12" Type I CentriFlow®	Cohesiveness:	Slight
Meter Capacity:	15 ft³/min	Density (lb/ft³):	140 - 150 lbs/ft³
Feed System:	Belt Conveyor	Inhibit Setting:	0.50%



Test #1	Mass Flow Rate = 20,000 - 25,000 lb/hr						
Run #	Actual Weight	Metered Weight	Metered/Actual	Delta Weight	% Error		
1	33.02	33.00	1.001	-0.02	-0.06%		
2	32.96	32.98	0.999	0.02	0.06%		
3	32.92	32.90	1.001	-0.02	-0.06%		
4	32.90	32.82	1.002	-0.08	-0.24%		
5	32.86	32.80	1.002	-0.06	-0.18%		
Average:			1.001				
STD:			0.00119				
% STD:			0.12%				
Additional Comments: Tested with a 12" CentriFlow® Type I Meter in the Reverse Flow Direction.							

### Accuracy Statement:

"The CentriFlow® Meter will provide accuracy to within  $\pm 0.25\%$  of reading when operating within  $\pm 10\%$  of the calibrated flow rate, as long as the flow rate is within the operational range of the meter."